



Intertek

CERTIFICATE OF CONFORMANCE

This authorizes the application of the ETL Verified Mark shown below to the models described in the Product Description section when made in accordance with the conditions set forth in the Verification Agreement and Qualification Testing Report.

Certificate Number: 3103597CRT-001-002

Applicant:

Agilent Technologies (S) Pte. Ltd.
No 1, Yishun Avenue 7
Singapore 768923

Manufacturer:

Agilent Technologies (S) Pte. Ltd.
No 1, Yishun Avenue 7
Singapore 768923

Applicant Contact: Mr. Edward Poon

Mfg. Contact: Mr. Edward Poon

Report No.: 3103597CRT-001

Product Description:

WireScope Pro Main unit (P/N: N2645A) and the WireScope Pro Dual remote unit (P/N: N2646A)
N2644A-100 - CAT-6A Universal Channel Smartprobe
N2644A-101 - CAT-6A Universal Link Smartprobe

This product has been tested and found to comply with the applicable electrical transmission characteristics specified in ANSI/TIA/EIA-568-B.1 Category 5 Level II, ANSI/TIA/EIA-B.2 Category 5e Level IIe, ANSI/TIA/EIA-568-B.2-1 Category 6 Level III (10/1995 Annex A & B), ISO/IEC 11801:2002, ISO/IEC 61935-1/Ed.2 (8-2005): Level IIe (Class D), Level III (Class E), Level IV (Class F).

This certificate, supported by your participation in the ETL Verification Program, is authorization to apply to your product or packaging the ETL Verification Mark.

Continuing compliance to this specification is monitored through production testing, onsite follow-up inspections by and Intertek Representative, and random sample testing.

Date ETL Verified: 10/6/2006

Revised Date: 7/12/2007

Certificate Issued By:



Kathy Heath, Program Administrator

This document is the property of Intertek and is not transferable. Only the Applicant may reproduce this document. The ETL Verified Mark may be applied only at the above noted location of the Party Authorized to Apply the Mark.

This document supersedes all previous ETL Verified Certificates of Conformance for the noted Certificate Number.