



Certificate of Calibration

ISO/IEC 17025:2017 and ANSI/NCSL Z540.1-1994

Certificate Number 1-11488382529-1

| | | | |
|----------------------------|--|--------------------------------|--|
| Model Number | 11667A | Customer | Keysight Technologies Korea Ltd |
| Manufacturer | Keysight Technologies Inc | | 25-12 Yeouido-dong |
| Description | DC-18 GHz power splitter, type N, 50 ohm | | Yeongdeungpo-gu |
| Serial Number | 07045 | | SEOUL 150-711 |
| Customer Asset No. | 11667A07045 | | Korea, Republic of |
| Options Installed | See Measurement Report | | |
| Date of Calibration | 12 Jul 2019 | Location of Calibration | Keysight Technologies Korea Ltd. |
| Procedure | STE-50114535-B.02.06 | | Singsong Center Bldg. #57, Yeouinaru-ro, |
| Temperature | (23 ± 5) °C | | Youngdeungpo-gu |
| Humidity | (50 ± 20) %RH | | Seoul 07327 |
| | | | KOREA, REPUBLIC OF |

This certifies that the equipment has been calibrated using applicable Keysight Technologies procedures and in compliance with ISO/IEC 17025:2005 and ANSI/NCSL Z540.1-1994 (R2002). The quality management system is registered to ISO 9001:2015. This report is NOT an accredited report by Korea Laboratory Accreditation Scheme, a ILAC MRA signatory.

As Received Conditions

The measured values of the equipment were observed in specification at the points tested. Additionally, the expanded measurement uncertainty intervals about the measured values were in specification.

Action Taken

- No corrective actions were necessary.

As Completed Conditions

The measured values of the equipment were observed in specification at the points tested. Additionally, the expanded measurement uncertainty intervals about the measured values were in specification.

Remarks or Special Requirements

This calibration report shall not be reproduced, except in full. The documented results relate to the equipment calibrated only.

The test limits stated in the report correspond to the published specifications of the equipment, at the points tested.

This calibration report may refer to equipment manufactured by HP, Agilent and Keysight as being manufactured by Keysight Technologies.

Based on the customer's request, the next calibration is due on 12 Jul 2020.

Keysight Technologies Korea Ltd.
Singsong Center Bldg. #57, Yeouinaru-ro,
Youngdeungpo-gu
Seoul 07327
KOREA, REPUBLIC OF

Kangouk Lee - Quality Manager

Certificate of Calibration

ISO/IEC 17025:2017 and ANSI/NCSL Z540.1-1994

Certificate Number 1-11488382529-1

Traceability Information

Technician ID Number N5241693

Measurements are traceable to the International System of Units (SI) via national metrology institutes (www.keysight.com/find/NMI) that are signatories to the CIPM Mutual Recognition Arrangement.

Calibration Equipment Used

| Model Number | Model Description | Equipment ID | Cal Due Date |
|--------------|---|--------------|--------------|
| 85054B | Standard mechanical calibration kit, DC to 18 GHz, type-N | 85054B00438 | 22 Nov 2020 |
| N5230A | PNA-L network analyzer | N5230A00214 | 13 Nov 2019 |

Traceability Table

| | Model | Model Description | Equipment ID | Certificate Number | Trace Value |
|-----|--------|---|--------------|------------------------------|--|
| W,R | 85054B | Standard mechanical calibration kit, DC to 18 GHz, type-N | 85054B00438 | 1-11253092369-1-ANAB:AC-1498 | Reflection Coefficient Transmission Coefficient |
| W | N5230A | PNA-L network analyzer | N5230A00214 | 1-10612086497-1 | |
| R | 85056A | Standard mechanical calibration kit, DC to 50 GHz, 2.4 mm | 85056A02325 | 1-10559247677-1-A2LA:2079.01 | Reflection Coefficient Transmission Coefficient |

Legend

W - Working Standard The calibration equipment used for the calibration of the Model indicated on the first page of the Certificate of calibration.

R - Reference Standard The Reference Standard (Accredited or NMI-calibrated ETE) used to provide traceability to the SI-Units for the calibration parameters listed.


Compliance with Specification

The uncertainty of measurement has been taken into account when determining compliance with specification, as per ILAC-G8:03/2009. If the expanded measurement uncertainty intervals centered about one or more measured values were both in as well as out of specification (upper or lower), it is not possible to state compliance or non-compliance based on a 95% coverage probability for the expanded measurement uncertainty.

An overall statement of compliance for all tests performed as received, and as completed (if any adjustments / repairs were performed) is included at the beginning of this report. Statements of compliance apply only to warranted specifications. When functional verification tests are performed, results are reported in the “Functional Test” section, and do not affect these statements of compliance. The status summaries relate to the tested item only. A final decision about whether the item's performance actually satisfies requirements of the user can only be made by the user.

Measurement results are reported as:

- Passed () - The measured values of the equipment were observed in specification at the points tested. Additionally, the expanded measurement uncertainty intervals about the measured values were in specification.
- Passed‡ (P‡) - The measured values of the equipment were observed in specification at the points tested. However, a portion of the expanded measurement uncertainty intervals about one or more measured values exceeded specification. Consequently, compliance with specification cannot be declared based on the stated coverage probability.
- Failed‡ (F‡) - One or more measured values of the equipment were observed out of specification at the points tested. However, a portion of the expanded measurement uncertainty intervals about one or more measured values were in specification. Consequently, non-compliance with specification cannot be declared based on the stated coverage probability.
- Failed (F) - One or more measured values of the equipment were observed out of specification at the points tested. Additionally, the expanded measurement uncertainty intervals about one or more measured values were entirely outside the specification.



() This result is indicated on the measurement report as a blank space in the column labeled “Status” or “Sts”.
 MU = 95% expanded measurement uncertainty.

Uncertainty of Measurement

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%. This probability corresponds to a coverage factor of k=2 for a normal distribution.



Certificate of Calibration

ISO/IEC 17025:2017 and ANSI/NCSL Z540.1-1994

Certificate Number 1-11488382529-1

Calibration Test Results Summary

| <u>Test Name</u> | <u>As Received Status</u> |
|-------------------------|---------------------------|
| INPUT REFL COEF | Passed |
| PORT2 INSERTION LOSS | Done |
| PORT3 INSERTION LOSS | Done |
| PORT2 to PORT3 TRACKING | Passed |
| EQUIVALENT SOURCE MATCH | Passed |

Tested Configuration

Tested Options 001
 (As Rec) 001

Model 11667A Serial 07045
 Options Tested 001

Test Date 12 Jul 2019
 Condition As Received

INPUT REFL COEF

Passed

The PHASE(deg) column is the measured PHASE.
 The MAG column is the measured MAGNITUDE.
 The MAXIMUM column is the spec for MAGNITUDE.
 The UNCERT column is the uncertainty for MAGNITUDE.

| TEST COND | PHASE | MAG | MAXIMUM | UNCERT. | Status |
|-----------|---------|-------|---------|---------|--------|
| 50 MHz | 84.4° | 0.001 | 0.070 | 0.0041 | |
| 100 MHz | 42° | 0.003 | 0.070 | 0.0041 | |
| 500 MHz | -24.5° | 0.005 | 0.070 | 0.0042 | |
| 1000 MHz | -114.7° | 0.003 | 0.070 | 0.0042 | |
| 1500 MHz | 134.8° | 0.002 | 0.070 | 0.0043 | |
| 2000 MHz | -0.3° | 0.005 | 0.070 | 0.0044 | |
| 2500 MHz | -89.9° | 0.008 | 0.070 | 0.0049 | |
| 3000 MHz | -174.5° | 0.013 | 0.070 | 0.0050 | |
| 3500 MHz | 121.9° | 0.018 | 0.070 | 0.0063 | |
| 4000 MHz | 75.5° | 0.023 | 0.070 | 0.0064 | |
| 4500 MHz | 29.2° | 0.028 | 0.111 | 0.0065 | |
| 5000 MHz | -24.8° | 0.029 | 0.111 | 0.0066 | |
| 5500 MHz | -86.3° | 0.024 | 0.111 | 0.0067 | |
| 6000 MHz | -160.1° | 0.015 | 0.111 | 0.0067 | |
| 6500 MHz | 109.8° | 0.016 | 0.111 | 0.022 | |
| 7000 MHz | 10.3° | 0.013 | 0.111 | 0.022 | |
| 7500 MHz | -99.6° | 0.005 | 0.111 | 0.022 | |
| 8000 MHz | 94.2° | 0.016 | 0.111 | 0.022 | |
| 8500 MHz | 28.4° | 0.032 | 0.184 | 0.023 | |
| 9000 MHz | -33.5° | 0.042 | 0.184 | 0.023 | |
| 9500 MHz | -94° | 0.040 | 0.184 | 0.023 | |
| 10000 MHz | -157.5° | 0.030 | 0.184 | 0.023 | |
| 10500 MHz | 142.3° | 0.018 | 0.184 | 0.023 | |
| 11000 MHz | 91° | 0.022 | 0.184 | 0.023 | |
| 11500 MHz | 19.5° | 0.037 | 0.184 | 0.023 | |
| 12000 MHz | -55.6° | 0.049 | 0.184 | 0.023 | |
| 12500 MHz | -128.6° | 0.050 | 0.184 | 0.023 | |
| 13000 MHz | 159.7° | 0.049 | 0.184 | 0.023 | |
| 13500 MHz | 88.2° | 0.052 | 0.184 | 0.024 | |
| 14000 MHz | 21.1° | 0.062 | 0.184 | 0.024 | |
| 14500 MHz | -56.5° | 0.056 | 0.184 | 0.024 | |
| 15000 MHz | -112.6° | 0.021 | 0.184 | 0.024 | |
| 15500 MHz | -80.7° | 0.030 | 0.184 | 0.024 | |
| 16000 MHz | -144.6° | 0.046 | 0.184 | 0.024 | |
| 16500 MHz | 148.2° | 0.035 | 0.184 | 0.024 | |
| 17000 MHz | 116.7° | 0.042 | 0.184 | 0.024 | |
| 17500 MHz | 52.4° | 0.077 | 0.184 | 0.024 | |
| 18000 MHz | -18.4° | 0.110 | 0.184 | 0.025 | |

The PHASE uncertainty is $\text{Arcsin}(\text{UNCERT}/\text{MAG})$ degree, except where MAG is less than UNCERT, in which case PHASE uncertainty is +/- 180 degree.

PORT2 INSERTION LOSS

Done

| <u>TEST COND</u> | <u>PHASE</u> | <u>MAG</u> | <u>UNCERT.</u> |
|-----------------------------|--------------|------------|----------------|
| NO SPECS FOR THIS PARAMETER | | | |
| 50 MHz | -5.2° | 6.03 dB | 0.040 dB |
| 100 MHz | -10.4° | 6.04 dB | 0.040 dB |
| 500 MHz | -51.6° | 6.06 dB | 0.058 dB |
| 1000 MHz | -103° | 6.08 dB | 0.043 dB |
| 1500 MHz | -154.5° | 6.11 dB | 0.043 dB |
| 2000 MHz | 154.1° | 6.14 dB | 0.043 dB |
| 2500 MHz | 102.8° | 6.18 dB | 0.067 dB |
| 3000 MHz | 51.6° | 6.22 dB | 0.067 dB |
| 3500 MHz | 0.4° | 6.25 dB | 0.078 dB |
| 4000 MHz | -50.8° | 6.25 dB | 0.078 dB |
| 4500 MHz | -102.1° | 6.27 dB | 0.078 dB |
| 5000 MHz | -153.4° | 6.29 dB | 0.079 dB |
| 5500 MHz | 155.2° | 6.32 dB | 0.079 dB |
| 6000 MHz | 103.8° | 6.38 dB | 0.079 dB |
| 6500 MHz | 52.9° | 6.46 dB | 0.22 dB |
| 7000 MHz | 2° | 6.48 dB | 0.22 dB |
| 7500 MHz | -48.8° | 6.47 dB | 0.22 dB |
| 8000 MHz | -99.7° | 6.43 dB | 0.22 dB |
| 8500 MHz | -150.9° | 6.38 dB | 0.24 dB |
| 9000 MHz | 157.8° | 6.30 dB | 0.24 dB |
| 9500 MHz | 106° | 6.21 dB | 0.24 dB |
| 10000 MHz | 54° | 6.22 dB | 0.24 dB |
| 10500 MHz | 2.1° | 6.31 dB | 0.24 dB |
| 11000 MHz | -48.9° | 6.40 dB | 0.24 dB |
| 11500 MHz | -99.8° | 6.40 dB | 0.24 dB |
| 12000 MHz | -150.9° | 6.33 dB | 0.24 dB |
| 12500 MHz | 157.5° | 6.23 dB | 0.24 dB |
| 13000 MHz | 105.8° | 6.19 dB | 0.24 dB |
| 13500 MHz | 53.9° | 6.07 dB | 0.24 dB |
| 14000 MHz | 1° | 6.01 dB | 0.24 dB |
| 14500 MHz | -51.5° | 6.10 dB | 0.24 dB |
| 15000 MHz | -103.8° | 6.19 dB | 0.25 dB |
| 15500 MHz | -155.7° | 6.34 dB | 0.25 dB |
| 16000 MHz | 153.5° | 6.40 dB | 0.25 dB |
| 16500 MHz | 101.2° | 6.17 dB | 0.25 dB |
| 17000 MHz | 47.8° | 6.26 dB | 0.25 dB |
| 17500 MHz | -5.3° | 6.46 dB | 0.25 dB |
| 18000 MHz | -56.9° | 6.81 dB | 0.26 dB |

PORT3 INSERTION LOSS

Done

| <u>TEST COND</u> | <u>PHASE</u> | <u>MAG</u> | <u>UNCERT.</u> |
|-----------------------------|--------------|------------|----------------|
| NO SPECS FOR THIS PARAMETER | | | |
| 50 MHz | -5.2° | 6.01 dB | 0.040 dB |
| 100 MHz | -10.4° | 6.02 dB | 0.040 dB |

Model 11667A Serial 07045
 Options Tested 001

Test Date 12 Jul 2019
 Condition As Received

PORT3 INSERTION LOSS (cont.)

| TEST COND | PHASE | MAG | UNCERT. |
|-----------|---------|---------|----------|
| 500 MHz | -51.5° | 6.04 dB | 0.058 dB |
| 1000 MHz | -102.9° | 6.06 dB | 0.043 dB |
| 1500 MHz | -154.2° | 6.08 dB | 0.043 dB |
| 2000 MHz | 154.4° | 6.10 dB | 0.043 dB |
| 2500 MHz | 103.1° | 6.13 dB | 0.067 dB |
| 3000 MHz | 51.9° | 6.16 dB | 0.067 dB |
| 3500 MHz | 0.8° | 6.19 dB | 0.078 dB |
| 4000 MHz | -50.5° | 6.20 dB | 0.078 dB |
| 4500 MHz | -101.7° | 6.23 dB | 0.078 dB |
| 5000 MHz | -153° | 6.25 dB | 0.078 dB |
| 5500 MHz | 155.7° | 6.29 dB | 0.079 dB |
| 6000 MHz | 104.5° | 6.37 dB | 0.079 dB |
| 6500 MHz | 53.8° | 6.43 dB | 0.22 dB |
| 7000 MHz | 2.8° | 6.44 dB | 0.22 dB |
| 7500 MHz | -48° | 6.45 dB | 0.22 dB |
| 8000 MHz | -98.9° | 6.42 dB | 0.22 dB |
| 8500 MHz | -149.9° | 6.36 dB | 0.24 dB |
| 9000 MHz | 158.9° | 6.29 dB | 0.24 dB |
| 9500 MHz | 107.1° | 6.19 dB | 0.24 dB |
| 10000 MHz | 55.1° | 6.20 dB | 0.24 dB |
| 10500 MHz | 3.4° | 6.31 dB | 0.24 dB |
| 11000 MHz | -47.5° | 6.40 dB | 0.24 dB |
| 11500 MHz | -98.3° | 6.39 dB | 0.24 dB |
| 12000 MHz | -149.3° | 6.31 dB | 0.24 dB |
| 12500 MHz | 159.1° | 6.21 dB | 0.24 dB |
| 13000 MHz | 107.5° | 6.17 dB | 0.24 dB |
| 13500 MHz | 55.8° | 6.04 dB | 0.24 dB |
| 14000 MHz | 2.9° | 5.98 dB | 0.24 dB |
| 14500 MHz | -49.5° | 6.04 dB | 0.24 dB |
| 15000 MHz | -102° | 6.15 dB | 0.25 dB |
| 15500 MHz | -153.7° | 6.32 dB | 0.25 dB |
| 16000 MHz | 155.6° | 6.37 dB | 0.25 dB |
| 16500 MHz | 103.3° | 6.15 dB | 0.25 dB |
| 17000 MHz | 50.1° | 6.26 dB | 0.25 dB |
| 17500 MHz | -2.6° | 6.49 dB | 0.25 dB |
| 18000 MHz | -53.7° | 6.70 dB | 0.26 dB |

PORT2 to PORT3 TRACKING

Passed

PORT2 is the right hand port and PORT3 is the left hand port, as seen reading the model number label.

The PHASE(deg) column is the measured PHASE.
 The MAG column is the measured MAGNITUDE.
 The MINIMUM and MAXIMUM columns are specs for MAGNITUDE.
 The UNCERT column is the uncertainty for MAGNITUDE.

| TEST COND | PHASE | MINIMUM | MAG | MAXIMUM | UNCERT. | Status |
|-----------|-------|-----------|-----------|----------|----------|--------|
| 50 MHz | 0° | -0.150 dB | -0.023 dB | 0.150 dB | 0.046 dB | |

PORT2 to PORT3 TRACKING (cont.)

| TEST COND | PHASE | MINIMUM | MAG | MAXIMUM | UNCERT. | Status |
|-----------|-------|-----------|-----------|----------|----------|--------|
| 100 MHz | 0° | -0.150 dB | -0.016 dB | 0.150 dB | 0.046 dB | |
| 500 MHz | 0° | -0.150 dB | -0.026 dB | 0.150 dB | 0.051 dB | |
| 1000 MHz | -0.1° | -0.150 dB | -0.022 dB | 0.150 dB | 0.049 dB | |
| 1500 MHz | -0.1° | -0.150 dB | -0.029 dB | 0.150 dB | 0.049 dB | |
| 2000 MHz | -0.1° | -0.150 dB | -0.023 dB | 0.150 dB | 0.050 dB | |
| 2500 MHz | -0.1° | -0.150 dB | -0.010 dB | 0.150 dB | 0.055 dB | |
| 3000 MHz | -0.2° | -0.150 dB | -0.019 dB | 0.150 dB | 0.053 dB | |
| 3500 MHz | -0.2° | -0.150 dB | -0.020 dB | 0.150 dB | 0.060 dB | |
| 4000 MHz | -0.3° | -0.150 dB | -0.020 dB | 0.150 dB | 0.056 dB | |
| 4500 MHz | -0.4° | -0.200 dB | -0.009 dB | 0.200 dB | 0.061 dB | |
| 5000 MHz | -0.4° | -0.200 dB | -0.011 dB | 0.200 dB | 0.063 dB | |
| 5500 MHz | -0.4° | -0.200 dB | -0.010 dB | 0.200 dB | 0.062 dB | |
| 6000 MHz | -0.6° | -0.200 dB | 0.007 dB | 0.200 dB | 0.072 dB | |
| 6500 MHz | -0.7° | -0.200 dB | 0.048 dB | 0.200 dB | 0.067 dB | |
| 7000 MHz | -0.7° | -0.200 dB | -0.001 dB | 0.200 dB | 0.079 dB | |
| 7500 MHz | -0.8° | -0.200 dB | 0.016 dB | 0.200 dB | 0.075 dB | |
| 8000 MHz | -1.1° | -0.200 dB | 0.024 dB | 0.200 dB | 0.078 dB | |
| 8500 MHz | -1.5° | -0.250 dB | 0.002 dB | 0.250 dB | 0.080 dB | |
| 9000 MHz | -1.6° | -0.250 dB | -0.076 dB | 0.250 dB | 0.071 dB | |
| 9500 MHz | -0.7° | -0.250 dB | -0.098 dB | 0.250 dB | 0.082 dB | |
| 10000 MHz | -0.9° | -0.250 dB | -0.014 dB | 0.250 dB | 0.074 dB | |
| 10500 MHz | -1.2° | -0.250 dB | 0.001 dB | 0.250 dB | 0.085 dB | |
| 11000 MHz | -1.4° | -0.250 dB | -0.006 dB | 0.250 dB | 0.083 dB | |
| 11500 MHz | -1.4° | -0.250 dB | -0.015 dB | 0.250 dB | 0.087 dB | |
| 12000 MHz | -1.5° | -0.250 dB | -0.023 dB | 0.250 dB | 0.095 dB | |
| 12500 MHz | -1.6° | -0.250 dB | -0.032 dB | 0.250 dB | 0.083 dB | |
| 13000 MHz | -1.5° | -0.250 dB | -0.032 dB | 0.250 dB | 0.092 dB | |
| 13500 MHz | -1.6° | -0.250 dB | -0.007 dB | 0.250 dB | 0.074 dB | |
| 14000 MHz | -1.6° | -0.250 dB | -0.002 dB | 0.250 dB | 0.082 dB | |
| 14500 MHz | -1.4° | -0.250 dB | 0.090 dB | 0.250 dB | 0.080 dB | |
| 15000 MHz | -2.3° | -0.250 dB | 0.078 dB | 0.250 dB | 0.079 dB | |
| 15500 MHz | -2.4° | -0.250 dB | 0.021 dB | 0.250 dB | 0.094 dB | |
| 16000 MHz | -2.5° | -0.250 dB | 0.004 dB | 0.250 dB | 0.091 dB | |
| 16500 MHz | -2.5° | -0.250 dB | 0.015 dB | 0.250 dB | 0.10 dB | |
| 17000 MHz | -2.5° | -0.250 dB | -0.003 dB | 0.250 dB | 0.099 dB | |
| 17500 MHz | -2.6° | -0.250 dB | 0.015 dB | 0.250 dB | 0.10 dB | |
| 18000 MHz | -3° | -0.250 dB | -0.087 dB | 0.250 dB | 0.12 dB | |

The PHASE uncertainty is $\text{Arcsin}(\text{Lin_Mu}/\text{Lin_Mag})$ degree, except where Lin_Mag is less than Lin_Mu , in which case PHASE uncertainty is +/- 180 degree.

Lin_Mu & Lin_Mag are linear mode of UNCERT & MAG, which can be calculated by:

$$\text{Lin_Mag} = 10^{(\text{MAG} / 20)}$$

$$\text{Lin_Mu} = 10^{(\text{UNCERT} / 20)}$$

EQUIVALENT SOURCE MATCH

Passed

PORT2 is the right hand port and PORT3 is the left hand port, as seen reading the model number label.

The PHASE(deg) column is the measured PHASE.
The MAG column is the measured MAGNITUDE.
The MAXIMUM column is the specs for MAGNITUDE.
The UNCERT column is the uncertainty for MAGNITUDE.

| TEST COND | PHASE | MAG | MAXIMUM | UNCERT. | Status |
|--------------------------------|---------|-------|---------|---------|--------|
| PORT2 EQUIVALENT SOURCE MATCH: | | | | | |
| 50 MHz | 163.6° | 0.003 | 0.048 | 0.0034 | |
| 100 MHz | 153° | 0.002 | 0.048 | 0.0036 | |
| 500 MHz | 83.5° | 0.001 | 0.048 | 0.0055 | |
| 1000 MHz | 103.8° | 0.002 | 0.048 | 0.0041 | |
| 1500 MHz | 102.8° | 0.003 | 0.048 | 0.0048 | |
| 2000 MHz | 86.5° | 0.005 | 0.048 | 0.0052 | |
| 2500 MHz | 60.3° | 0.008 | 0.048 | 0.0042 | |
| 3000 MHz | 41.3° | 0.009 | 0.048 | 0.0063 | |
| 3500 MHz | 13.1° | 0.009 | 0.048 | 0.0041 | |
| 4000 MHz | -18.3° | 0.010 | 0.048 | 0.0062 | |
| 4500 MHz | -57° | 0.011 | 0.091 | 0.0054 | |
| 5000 MHz | -86.3° | 0.015 | 0.091 | 0.0052 | |
| 5500 MHz | -110.3° | 0.020 | 0.091 | 0.0066 | |
| 6000 MHz | -127.2° | 0.024 | 0.091 | 0.0045 | |
| 6500 MHz | -131.4° | 0.028 | 0.091 | 0.0072 | |
| 7000 MHz | -156° | 0.040 | 0.091 | 0.0051 | |
| 7500 MHz | -170° | 0.042 | 0.091 | 0.0070 | |
| 8000 MHz | 176.1° | 0.044 | 0.091 | 0.0064 | |
| 8500 MHz | 159.7° | 0.044 | 0.142 | 0.0061 | |
| 9000 MHz | 140.1° | 0.041 | 0.142 | 0.0075 | |
| 9500 MHz | 117.8° | 0.033 | 0.142 | 0.0052 | |
| 10000 MHz | 81.1° | 0.023 | 0.142 | 0.0083 | |
| 10500 MHz | 24.1° | 0.019 | 0.142 | 0.0059 | |
| 11000 MHz | -30.2° | 0.028 | 0.142 | 0.0077 | |
| 11500 MHz | -62.4° | 0.042 | 0.142 | 0.0074 | |
| 12000 MHz | -86.5° | 0.056 | 0.142 | 0.0068 | |
| 12500 MHz | -105.6° | 0.066 | 0.142 | 0.0088 | |
| 13000 MHz | -123.7° | 0.072 | 0.142 | 0.0057 | |
| 13500 MHz | -140° | 0.076 | 0.142 | 0.0088 | |
| 14000 MHz | -156° | 0.071 | 0.142 | 0.0064 | |
| 14500 MHz | -171.4° | 0.065 | 0.142 | 0.0081 | |
| 15000 MHz | 173° | 0.058 | 0.142 | 0.0085 | |
| 15500 MHz | 153.4° | 0.051 | 0.142 | 0.0068 | |
| 16000 MHz | 130.9° | 0.045 | 0.142 | 0.011 | |
| 16500 MHz | 104.6° | 0.048 | 0.142 | 0.0067 | |
| 17000 MHz | 76.1° | 0.057 | 0.142 | 0.010 | |
| 17500 MHz | 45.6° | 0.070 | 0.142 | 0.0088 | |
| 18000 MHz | 28.2° | 0.081 | 0.142 | 0.0099 | |

Model 11667A Serial 07045
 Options Tested 001

Test Date 12 Jul 2019
 Condition As Received

EQUIVALENT SOURCE MATCH (cont.)

| TEST COND | PHASE | MAG | MAXIMUM | UNCERT. | Status |
|--------------------------------|---------|-------|---------|---------|--------|
| PORT3 EQUIVALENT SOURCE MATCH: | | | | | |
| 50 MHz | 172.9° | 0.005 | 0.048 | 0.0034 | |
| 100 MHz | 163.8° | 0.005 | 0.048 | 0.0036 | |
| 500 MHz | 107.8° | 0.003 | 0.048 | 0.0055 | |
| 1000 MHz | 111° | 0.006 | 0.048 | 0.0042 | |
| 1500 MHz | 81.3° | 0.005 | 0.048 | 0.0049 | |
| 2000 MHz | 82.2° | 0.008 | 0.048 | 0.0053 | |
| 2500 MHz | 60.3° | 0.009 | 0.048 | 0.0043 | |
| 3000 MHz | 48.8° | 0.009 | 0.048 | 0.0064 | |
| 3500 MHz | 31.6° | 0.008 | 0.048 | 0.0042 | |
| 4000 MHz | 9.4° | 0.006 | 0.048 | 0.0061 | |
| 4500 MHz | -28.5° | 0.006 | 0.091 | 0.0055 | |
| 5000 MHz | -60.9° | 0.007 | 0.091 | 0.0053 | |
| 5500 MHz | -86.2° | 0.009 | 0.091 | 0.0066 | |
| 6000 MHz | -100.9° | 0.015 | 0.091 | 0.0045 | |
| 6500 MHz | -111.6° | 0.025 | 0.091 | 0.0071 | |
| 7000 MHz | -133.5° | 0.030 | 0.091 | 0.0050 | |
| 7500 MHz | -146.3° | 0.035 | 0.091 | 0.0069 | |
| 8000 MHz | -158.6° | 0.043 | 0.091 | 0.0063 | |
| 8500 MHz | -171.9° | 0.045 | 0.142 | 0.0061 | |
| 9000 MHz | 170.7° | 0.057 | 0.142 | 0.0074 | |
| 9500 MHz | 131.1° | 0.052 | 0.142 | 0.0052 | |
| 10000 MHz | 97° | 0.033 | 0.142 | 0.0083 | |
| 10500 MHz | 48.7° | 0.022 | 0.142 | 0.0058 | |
| 11000 MHz | -12.5° | 0.024 | 0.142 | 0.0076 | |
| 11500 MHz | -52.9° | 0.036 | 0.142 | 0.0073 | |
| 12000 MHz | -78° | 0.049 | 0.142 | 0.0068 | |
| 12500 MHz | -97.7° | 0.060 | 0.142 | 0.0088 | |
| 13000 MHz | -114.6° | 0.066 | 0.142 | 0.0056 | |
| 13500 MHz | -129.6° | 0.066 | 0.142 | 0.0088 | |
| 14000 MHz | -143.5° | 0.070 | 0.142 | 0.0065 | |
| 14500 MHz | -176.4° | 0.050 | 0.142 | 0.0081 | |
| 15000 MHz | -165.1° | 0.037 | 0.142 | 0.0086 | |
| 15500 MHz | -174.4° | 0.035 | 0.142 | 0.0067 | |
| 16000 MHz | 171.4° | 0.032 | 0.142 | 0.011 | |
| 16500 MHz | 151.2° | 0.029 | 0.142 | 0.0065 | |
| 17000 MHz | 109° | 0.048 | 0.142 | 0.010 | |
| 17500 MHz | 66.6° | 0.061 | 0.142 | 0.0085 | |
| 18000 MHz | 28.7° | 0.078 | 0.142 | 0.0097 | |

The PHASE uncertainty is $\text{Arcsin}(\text{UNCERT}/\text{MAG})$ degree, except where MAG is less than UNCERT, in which case PHASE uncertainty is +/- 180 degree.