

Keysight 16380C Capacitance Standard Set

Notices

© Keysight Technologies
1985 - 2019

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Keysight Technologies, Inc. as governed by United States and international copyright laws.

Trademark Acknowledgments

Manual Part Number

16380-90221

Edition

Edition 7, August 20, 2019

Printed in Malaysia

Published by:

Keysight Technologies International
Japan G.K,
1-3-3 Higashikawasaki-cho
Chuo-ku
Kobe-shi, Hyogo, Japan

Warranty

THE MATERIAL CONTAINED IN THIS DOCUMENT IS PROVIDED "AS IS," AND IS SUBJECT TO BEING CHANGED, WITHOUT NOTICE, IN FUTURE EDITIONS. FURTHER, TO THE MAXIMUM EXTENT PERMITTED BY APPLICABLE LAW, KEYSIGHT DISCLAIMS ALL WARRANTIES, EITHER EXPRESS OR IMPLIED WITH REGARD TO THIS MANUAL AND ANY INFORMATION CONTAINED HEREIN, INCLUDING BUT NOT LIMITED TO THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. KEYSIGHT SHALL NOT BE LIABLE FOR ERRORS OR FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH THE FURNISHING, USE, OR PERFORMANCE OF THIS DOCUMENT OR ANY INFORMATION CONTAINED HEREIN. SHOULD KEYSIGHT AND THE USER HAVE A SEPARATE WRITTEN AGREEMENT WITH WARRANTY TERMS COVERING THE MATERIAL IN THIS DOCUMENT THAT CON-

FLICT WITH THESE TERMS, THE WARRANTY TERMS IN THE SEPARATE AGREEMENT WILL CONTROL.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Declaration of Conformity

Declarations of Conformity for this product and for other Keysight products may be downloaded from the Web. Go to <http://www.keysight.com/go/conformity>. You can then search by product number to find the latest Declaration of Conformity.

U.S. Government Rights

The Software is "commercial computer software," as defined by Federal Acquisition Regulation ("FAR") 2.101. Pursuant to FAR 12.212 and 27.405-3 and Department of Defense FAR Supplement ("DFARS") 227.7202, the U.S. government acquires commercial computer software under the same terms by which the software is customarily provided to the public. Accordingly, Keysight provides the Software to U.S. government customers under its standard commercial license, which is embodied in its End User License Agreement (EULA), a copy of which can be found at <http://www.keysight.com/find/sweula>. The license set forth in the EULA represents the exclusive authority by which the U.S. government may use, modify, distribute, or disclose the Software. The EULA and the license set forth therein, does not require or permit, among other things, that Keysight: (1) Furnish technical information related to commercial computer software or commercial computer software documentation that is not customarily provided to the public; or (2) Relinquish to, or otherwise provide, the government rights in excess of these rights customarily provided to the public to use, modify, reproduce, release, perform, display, or disclose commercial computer software or commercial computer software documentation. No addi-

tional government requirements beyond those set forth in the EULA shall apply, except to the extent that those terms, rights, or licenses are explicitly required from all providers of commercial computer software pursuant to the FAR and the DFARS and are set forth specifically in writing elsewhere in the EULA. Keysight shall be under no obligation to update, revise or otherwise modify the Software. With respect to any technical data as defined by FAR 2.101, pursuant to FAR 12.211 and 27.404.2 and DFARS 227.7102, the U.S. government acquires no greater than Limited Rights as defined in FAR 27.401 or DFARS 227.7103-5 (c), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.

1. General Information	
Description	5
Contents	5
Initial Inspection	6
2. Specifications and Supplemental Performance Characteristics	
Specifications	7
Operating Conditions	7
Dimensions	7
Weight	7
Accessories Furnished	8
Supplemental Performance Characteristics	8
Capacitance Temperature Coefficient	8
Capacitance Stability	8
Maximum Allowable Voltage/Current	8
Residual Capacitance	8
Storage Conditions	9
3. Service	
Calibration	11
Recommended Calibration Cycle	11
Repair	11

1 General Information

Description

The Keysight 16380C Capacitance Standard Set consists of three precision capacitors 0.01 μF (16385A), 0.1 μF (16386A), and 1 μF (16387A). One precision capacitor 10 μF (16388A) is available with Option 001. All four capacitors have high capacitance stability and are virtually unaffected by changes in environmental temperature, ensuring measurement repeatability and reliability.

The 16380C was designed for use in calibrating precision impedance measuring instruments (such as LCR meters, impedance analyzers, etc.) that have capacitance ranges between 0.01 μF and 10 μF . The 16380C can be used to verify capacitance measurement accuracy over the 0.01 μF to 10 μF range.

Contents

The 16380C contents are listed in [Table 1-1](#).

Table 1-1

Description	Keysight Part No.	Qty.
16385A (0.01 μF)	Not Assigned	1
16386A (0.1 μF)	Not Assigned	1
16387A (1 μF)	Not Assigned	1
16388A (10 μF) (Option 001)	Not Assigned	1
BNC(f)-(f) Adapters	1250-0080	4
Case	16380-85104	1
Calibration Report	Not Assigned	1
Operating Note (This Manual)	16380-90221	1

Initial Inspection

Inspect the shipping container for damage. If the shipping container or cushioning material is damaged, it should be kept until the contents of the shipment have been checked for completeness and the instrument has been checked mechanically. The contents of the shipment should be as shown in **Table 1-1**. If the shipment is incomplete, or if there is mechanical damage or defects, notify the nearest Keysight office. If the shipping container is damaged, or the cushioning material shows signs of stress, notify the carrier as well as the Keysight office. Keep the shipping materials for carrier's inspection. The Keysight office will arrange for repair or replacement at Keysight Option, without waiting for claim settlement.

2 Specifications and Supplemental Performance Characteristics

Specifications

Table 2-1

Model	16385A	16386A	16387A	16388A (opt.001)
Nominal Value ¹	0.01 μ F	0.1 μ F	1 μ F	10 μ F
Nominal Accuracy ¹		$\pm 0.1\%$		$\pm 0.05\%$
Dissipation Factor ¹	$\leq 4 \times 10^{-4}$	$\leq 5 \times 10^{-4}$	$\leq 7 \times 10^{-4}$	$\leq 5 \times 10^{-4}$

1. Specified at 1 kHz under the following ambient conditions:
Temperature: 23 ± 5 °C

Operating Conditions

Ambient Temperature: 23 ± 5 °C

Relative Humidity: Less than 70% RH

Dimensions

16385A, 16386A, 16387A, and 16388A (opt.001): 142(W) \times 88(D) \times 112(H) mm

Carrying Case: 355(W) \times 340(D) \times 170(H) mm

Weight

16385A, 16386A, and 16387A: approximately 1.3 kg each

16388A (opt.001): approximately 1.7 kg

Total: approximately 7.0 kg¹/approximately 8.7 kg²

1. Includes three capacitance standards, a carrying case, and four BNC adapters.
2. Includes four capacitance standards, a carrying case, and four BNC adapters.

Accessories Furnished

- BNC (f)-(f) Adapters, 4ea: PN 1250-0080
- Carrying Case, 1 ea.: PN 16380-85104
- Operating Note, 1 ea.: PN 16380-90221

Supplemental Performance Characteristics

Capacitance Temperature Coefficient

16385A, 16386A, 16387A, and 16388A (opt.001): -20 to 35 ppm/°C

Capacitance Stability

16385A, 16386A, 16387A, and 16388A (opt.001): <50ppm/year @ 1 kHz, 23±5 °C

Maximum Allowable Voltage/Current

16385A, 16386A, 16387A, and 16388A (opt.001): 40 V peak (ac+dc)/
 100 mA rms


Residual Capacitance

Table 2-2

	16385A	16386A	16387A	16388A (opt.001)
C_{HG}^1	<20 pF	<20 pF	<35 pF	<50 pF
C_{LG}^1	<20 pF	<20 pF	<35 pF	<50 pF

1. See the figure below.

Figure 2-1



Specifications and Supplemental Performance Characteristics
Supplemental Performance Characteristics

C_{HL} : 4 terminal capacitance

C_{HG} , C_{LG} : Stray capacitance to guard

Storage Conditions

Temperature: $-10\text{ }^{\circ}\text{C}$ to $+55\text{ }^{\circ}\text{C}$

Relative Humidity: Less than 85% RH at $40\text{ }^{\circ}\text{C}$

Specifications and Supplemental Performance Characteristics
Supplemental Performance Characteristics

3 Service

Calibration

Keysight will calibrate the 16380C. For complete information (price, time required, etc.) on how to have the 16380C calibrated, contact the nearest Keysight Sales and Service Office.

Recommended Calibration Cycle

The 16380C should be calibrated at least once a year. More frequent calibration may be required if the 16380C is used in very hot or very cold environments.

Repair

The 16380C contains no replaceable components. If one of the capacitors becomes damaged, or if its capacitance value is outside specified Limits, the capacitor must be replaced. For complete information on service, contact the nearest Keysight Sales and Service Office.

Service
Repair

This information is subject to change without notice.
© Keysight Technologies 1985 - 2019
Edition 7, August 20, 2019



16380-90221
www.keysight.com