

Keysight X-Series Signal Analyzers Multi-Touch

Option ESC External Source Control

Notices

© Copyright 2016 Keysight Technologies, Inc.

The information contained in this document is subject to change without notice.

Keysight Technologies makes no warranty of any kind with regard to this material, including but not limited to, the implied warranties of merchantability and fitness for a particular purpose.

Keysight Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Manual Part Number

N9020-90262

Edition

Edition 2, February 2016

Printed in USA/Malaysia

Published by:

Keysight Technologies, Inc.
1400 Fountaingrove Parkway
Santa Rosa, CA 95403

Option ESC - External Source Control Installation Kit

Products Affected:	N9000B CXA Signal Analyzer N9010B EXA Signal Analyzer N9020B MXA Signal Analyzer N9030B PXA Signal Analyzer N9040B UXA Signal Analyzer
Serial Numbers:	All
To Be Performed By:	(X) Keysight Service Center (X) Personnel Qualified by Keysight (X) Customer
Estimated Installation Time:	0.5 Hours

Introduction

This kit contains all of the instructions required to install Option ESC on a Keysight CXA, EXA, MXA, PXA, or UXA Signal Analyzer. No tools are required.

The N9040B UXA requires software version A.17.04 or later.

The latest revision of the X-Series Signal Analyzers software may be downloaded from

http://www.keysight.com/find/xseries_software.

This option is licensed for one instrument model/serial number combination. The license key will only install on the designated instrument.

Contents

Quantity	Description
1	Installation Note
1	Option Upgrade Entitlement Certificate
3	Cable Assembly - RG/233 BNC - BNC 48 in.
1	Cable Assembly - LAN Crossover

Tools Required

- Personal computer with internet access and USB port
- USB storage device with > 2 GB free memory

NOTE

If this upgrade is being performed at an Keysight Service Center, be sure to send the included cables when the instrument is returned to its owner. The cables are used to connect the signal source to the signal analyzer.

Installation Procedure over USB

1. Locate the Option Upgrade Entitlement Certificate (5964-5178) from the kit.
2. Redeem the Option Upgrade Entitlement Certificate by following the instructions on the Certificate.
3. After redeeming your Option Upgrade Entitlement Certificate you will receive an email with an attached License File.
4. Locate a USB storage device. Perform a virus scan on this device before use.
5. Save the License File to the root directory of the USB storage device.
6. Connect the USB storage device to the signal analyzer USB port. Windows will detect the new hardware and may display the configuration menu shown in **Figure 1**. This menu may be configured according to your preferences.

Figure 1 USB Storage Device Configuration Menu



7. The signal analyzer will automatically consume the License File. (This may take a few minutes) When the License File is consumed the Keysight License Manager will display a “Successful License Installation” message similar to the one shown in [Figure 2](#).

Figure 2 Successful License Installation



Alternate Installation Procedure

The License File can be manually installed over USB or LAN by placing the license file in the following folder on the signal analyzer

C:\Program Files\Agilent\licensing

Verify the Installation and Instrument Software Version

1. Cycle power on the analyzer and wait for the analyzer to boot-up completely.
2. Press **System, Show System** to display a list of installed options.
3. Verify that the newly installed option number appears on the list.

Option ESC - External Source Control Installation Kit

Utilities, Adjustments, and Performance Verification Tests

Utilities Required

None

Adjustments Required

None

Performance Testing Required

None

For assistance, contact your nearest Keysight Technologies Sales and Service Office. To find your local Keysight office access the following URL:

<http://www.keysight.com/find/assist>



This information is subject to change without notice.

© Keysight Technologies 2016

Edition 2, February 2016

N9020-90262

www.keysight.com